



DECLARATION OF CONFORMITY

Manufacturer:

LIW Care Technology Sp. z o.o.
ul. Golfowa 7
94-406 Łódź

Hereby declares, that:

SPECIAL PUSHCHAIR FOR CHILDREN **RICO BUGGY** **Sizes: 1,2**

bearing CE mark, is a medical device of I Class, Rule 1 in accordance with Annex 9 of European Medical Devices Directive 93/42/EEC and meets the essential requirements from Appendix 1 of this Directive.

The conformity assessment was carried out on the basis of Annex 7 of Council Directive 93/42 / EEC.

The product meets the requirements of the Act of 20 May 2010 on medical devices (Journal of Laws 2017 item 211), Regulation of the Minister of Health of February 17, 2016. on essential requirements and conformity assessment procedures for medical devices (Journal of Laws of 2016, item 211).

Medical device is in conformity with the following harmonized standards:

- **PN-EN 12182:2012**
- **PN-EN ISO 14971:2012**
- **PN-EN 1041+A1:2013-12**
- **PN-EN ISO 15223-1:2017-02**

On behalf of the manufacturer:
LIW Care Technology Sp. z o.o.

Prezes Zarządu

Tomasz Chmielecki

Tomasz Chmielecki
PREZES ZARZĄDU

Łódź, 01.02.2019 r.

LIW CARE TECHNOLOGY Sp. z o.o.
94-406 Łódź, ul. Golfowa 7
NIP: 729-266-53-87, REG. 100715121
KRS:0000333719